



# EVALUATION OF THE S.T.A.R.<sup>®</sup> TOTAL ANKLE REPLACEMENT: *AN EVOLUTION IN DESIGN*

Orthopaedic Research Laboratories  
Lutheran Hospital  
a Cleveland Clinic hospital  
Cleveland, Ohio

Paul D. Postak, B.Sc.  
Mircea Rosca, M.S.M.E.  
A. Seth Greenwald, D.Phil.(Oxon)

## INTRODUCTION

The use of total ankle replacements is increasing as an alternative to arthrodesis that allows patients to maintain both mobility and function concomitant with pain relief. Their success, however, remains mixed mostly attributed to the interaction of design geometries and the complexity of ankle joint motion.<sup>5</sup> This has been exemplified by failures of designs with excessive constraint or insufficient material durability leading to loosening and wear.<sup>6,9,14-16</sup> The expectation of these failures makes paramount the evaluation of anticipated device performance, particularly long-term fatigue and wear characteristics, prior to widespread clinical use.

This exhibit describes the wear characteristics of a three-part mobile total ankle replacement during 10 to 20 years of simulated walking in a novel, multi-axial gait simulator (US Patent Application Serial No. 11/498,313). These results are useful in demonstrating safety and effectiveness of this device and also present a pre-clinical evaluation methodology for future ankle designs.

## THE DESIGN

Testing was performed on the S.T.A.R.<sup>®</sup> mobile ankle device (Scandinavian Total Ankle Replacement, Link Orthopaedics, Rockaway, NJ, USA). The S.T.A.R.<sup>®</sup> is a three-part ankle replacement consisting of an ultra-high molecular weight polyethylene (UHMWPE) mobile bearing positioned between highly polished cobalt chrome molybdenum (CoCrMo) talar and tibial components. The device relies on biological ingrowth for fixation and purports minimal bone resection. The mobile bearing is constrained to only dorsi and plantar flexion motion on the talar articulation, while allowing both translation and rotation to occur in a transverse plane on the tibial articulation. The indications for use are treatment of the painful arthritic ankle as an alternative to arthrodesis. This device has been implanted outside the United States since 1990 but has only recently received the recommendation of the Food and Drug Administration (FDA) Orthopaedic Advisory Panel for use in the United States based on both a clinical trial and the information presented in this study.




Figure 1: S.T.A.R.<sup>®</sup> Mobile Ankle Replacement.

## THE SIMULATION

Ankle motion simulation poses several technical difficulties including a limited knowledge of ankle loads and motions. Data from several researchers<sup>1,10,13</sup> was synthesized into a walking profile. Specimens were aligned in 5 degrees of plantar flexion and loaded statically to 3000 N of joint compression. Under displacement control, the prosthesis was articulated through 15 degrees of fully reversing flexion ( $R = -2$ ) and 2.0 degrees of fully reversing axial rotation ( $R = -1$ ). In addition, the prosthesis was articulated 2.5 mm of fully reversing anterior translation ( $R = -1$ ). All motions were sinusoidal at a rate of 1 Hz. Internal rotation was phase shifted -90 degrees with respect to dorsi/plantar flexion and anterior translation.

Ankle Simulator Wear Motion Profile



Five specimens ( $n=5$ ) were tested in a series of identical, custom, single-station motion simulators for 10 million motion cycles (Mc). All tests were conducted in bovine calf serum lubricant diluted to a total protein concentration of 20 grams per liter with phosphate buffered saline as specified in the ASTM guide F1714-96. Additionally, the lubricant contained 0.2% sodium azide and ethylene-diaminetetraacetic acid (EDTA) at a concentration of 20 mM.




Figure 2: Multi-axial gait simulator.


## DETERMINING WEAR

Wear was determined by the gravimetric weight loss method described in ASTM F1714. Specimens were evaluated before testing, every 500,000 cycles for 1 million cycles, and every 1 million cycles thereafter. Individual components were weighed on a Sartorius ME235S analytical balance (Sartorius AG, Goettingen, Germany) with a resolution of 0.01 mg. At each of the test intervals fresh lubricant was replaced and the wear particles analyzed. The cumulative weight loss and wear rate were reported for each specimen at each interval after correcting for fluid absorption in two load soak controls (n=2).

## RESULTS

Mean volumetric wear rate for the UHMWPE Core component over 10 million cycles was 5.7 mm<sup>3</sup> per million cycles.

Average loss of height in the UHMWPE Core component was 0.19 mm on the talar articular surface and 0.18 mm on the tibial articular surface.



The inferior articular surface of the S.T.A.R.® Core component after 10 million cycles demonstrating extensive abrasive polishing and minor pitting.



Figure 3: Specimen STAR-04 after 10 Mc.

## DISCUSSION

Trauma, most often, is the antecedent event giving rise to ankle degenerative joint disease (DJD) where impairment of the subtalar complex is not always appreciated. Early ankle designs provided excessive geometric constraint across the tibial-talar complex not fully accounting for the subtalar role in assisting ankle rotation. Fixation failure was all too common with failure rates ranging between 52% and 95% at 10 years.<sup>6,9,14-16</sup> Mobile bearing ankle designs seek to accommodate rotation at the tibial-talar surface.<sup>3,12</sup>

This study demonstrates that within a 10-year period of simulated function, the volumetric wear debris generated is substantially less than what has been reported in previous hip and knee simulator reports.<sup>2,4</sup> Although variance in simulator and prosthetic design can account for data differences, the findings support a reduced UHMWPE particle load as well as a reduction in wear rate employing conventional UHMWPE.

Ankle replacement to-date has seen limited clinical use by comparison to arthrodesis. Its success is seen to be dependent not only on design, but on careful patient selection, precise technical placement, correction of deformity, and the ability to realize sound fixation over time.<sup>7,8,11</sup> The relief of pain and return of function, as in any device, are the clinical endpoints. Mobility in total ankle design offers an appreciation of joint biomechanics and suggests that the above clinical goals are achievable.

## CONCLUSION

The S.T.A.R.® mobile bearing ankle device represents a design which accommodates the need to achieve rotation at the tibial-talar articulation and overcomes the geometric constraint of earlier two-part ankle systems.

The simulation suggests that the UHMWPE bearing exhibits durability over long-term dynamic function and offers reduction in both volumetric wear and wear rate by comparison to hip and knee simulator reports.

These laboratory simulator evaluations assist an understanding of anticipated performance of total ankle designs particular to their material wear characteristics, and should be required of all future designs to assure safety and effectiveness.

## REFERENCES

1. Bell CJ, Parsons P, Fisher J: Simulation of polyethylene wear in ankle joint prostheses. *Transactions of the Orthopaedic Research Society* 30:246, 2005.
2. Benjamin J, Szivek J, Dersam G, Persselin S, Johnson R: Linear and volumetric wear of tibial inserts in posterior cruciate-retaining knee arthroplasties. *Clin Orthop Relat Res* 392:131-138, 2001.
3. Buechel FF, Pappas MJ, Iorio LJ: New Jersey low contact stress total ankle replacement: biomechanical rationale and review of 23 cementless cases. *Foot Ankle* 8(6):279-290, 1988.
4. Engh CA, Sychterz CJ, Engh CA Jr.: Conventional ultra-high molecular weight polyethylene: a gold standard of sorts. *Instr Course Lect* 54:183-187, 2005.
5. Inman VT: *The Joints of the Ankle*. Williams & Wilkins, Baltimore, MD, 1976.
6. Kitaoka HB, Patzer GL: Clinical results of the Mayo total ankle arthroplasty. *J Bone Joint Surg* 78A(11):1658-1664, 1996.
7. Kofoed H: Cylindrical cemented ankle arthroplasty: a prospective series with long term follow up. *Foot Ankle Int* 16(8):474-479, 1995.
8. Kofoed H, Sorensen TS: Ankle arthroplasty for rheumatoid arthritis and osteoarthritis: prospective long term study of cemented replacements. *J Bone Joint Surg* 80B(2):328-332, 1998.
9. Matejczyk MB, Greenwald AS, Black JD: Ankle implant systems: laboratory evaluation and clinical correlation. *Transactions of the Orthopaedic Research Society* 4:114, 1979.
10. Proctor P, Paul JP: Ankle Joint Biomechanics. *J Biomechanics* 15(9):627-634, 1982.
11. Pyevich MT, Alvine FG, Saltzman CL, Callaghan JJ: Total ankle arthroplasty: 2-11 year follow up of a unique design. Presented at the 64th Annual Meeting of the American Academy of Orthopaedic Surgeons, 1997.
12. Raikin SM, Heim CS, Plaxton NA, Greenwald AS: Mobility characteristics of total ankle replacements. *Proceedings of the American Academy of Orthopaedic Surgeons* 1:617, 2000.
13. Stauffer RN, Chao, EYS, Brewster, RC: Force and motion analysis of the normal, diseased, and prosthetic ankle joint. *Clin Orthop Relat Res* 127:189, 1977.
14. Takakura Y, Tanaka Y, Tamai S, Masuhara K: Ankle arthroplasty. A comparative study of cemented metal and uncemented ceramic prostheses. *Clin Orthop Relat Res* 252:209-216, 1990.
15. Unger AS, Cluglis AE, Mow CS, Figgie HE: Total ankle arthroplasty in rheumatoid arthritis: a long term follow up study. *Foot Ankle* 8:173-177, 1988.
16. Wynn AH, Wilde AH: Long term follow up of the Conaxial (Beck-Steffee) total ankle arthroplasty. *Foot Ankle* 13(6):303-306, 1992.